

Peptide Standard Set Protocol and Product Information Sheet

Product Category:	Mass Spectrometry Standards
Catalog Number(s):	s6104-100ug
Product Name:	Peptide Calibration Standard Set
Alternative Name(s):	Peptide Standard Mixture
Contains (100 µg of each):	Leu-Enkephalin, Gonadoliberin, Angiotensin I, and Neurotensin
CAS Number:	Combination of Compounds
Molecular Weight:	555.6, 1181.6, 1295.7, 1691.0
Monoisotopic Masses [M + H] ⁺ :	556.277, 1182.581, 1296.685, 1690.928* *(-H ₂ O is 1672.917)
Storage:	-20°C or below

**The masses listed are the expected masses, instrument settings, solvent purity, and other conditions can affect the abundance of the [M+H]⁺ ion abundance. Other common ions include the [M+Na]⁺, [M+H₂O]⁺, [M+K]⁺, etc. In these cases add the appropriate mass to theoretical monoisotopic mass to identify peptide: 22, 18, 38 respectively. Loss of water (-18) is also a common occurrence (especially if ionization energy is elevated).*

Peptide standards are very useful for calibrating and verifying MALDI-MS and LCMS instrumentation. These standards are generally used to confirm that instrumentation is properly calibrated and to perform regular check-ups and validation of instrumentation. These standards are soluble in water, as well as acetonitrile/water co-solvents.

The procedure below is intended to be a general protocol or a starting point, not necessarily the best for your particular application.

Standard Preparation (for MALDI-MS)

1. Dissolve the contents of the tube in 500 µL of Proteomics Grade water or equivalent. Allow the standard(s) to sit for 2-5 minutes, then vortex moderately. This gives 200 µg/mL solution of each peptide standard.
2. Prepare a 5 mg/mL CHCA stock solution in 50% Acetonitrile:Water with 0.1% TFA.
3. Combine 5 µL of the peptide standard stock solution (prepared in step 1) with 95 µL CHCA solution (prepared in step 2). Vortex standard:matrix solution moderately. *Other matrices and/or solvent compositions can be used.*
4. Apply 0.2 to 1.0 µL of this solution onto the MALDI sample plate.
5. Allow the matrix:sample to co-crystallize through evaporation at room temperature.
6. Place MALDI plate in MALDI-MS Ion Source and analyze samples.

Thin Layer Method is also a good option, although this is not covered in this product sheet.